

U-CyTech BV

Yalelaan 48
3584 CM Utrecht
The Netherlands
P +31.30.253 5960
F +31.30.253 9344
INFO@ucytech.com
www.ucytech.com

Data sheet Monkey IL-12/23p40 ELISPOT antibody pair; 10-plate format

Cat. No.: CT619-10

Coating antibodies (2 vials)

Product: Monoclonal antibody to monkey interleukin 12/23p40 (IL-12/23p40) (binds both IL-12 (p35/p40) and IL-23 (p19/p40) heterodimers via the p40 subunit)

Isotype: Mouse IgG₁

Production: *In vitro* using serum free medium

Purification: Ion exchange chromatography

Contents: Each vial contains sufficient material for coating of five 96-well ELISPOT plates

Buffer: Prior to lyophilization: 0.25 ml PBS + 125 mM trehalose

Application: Coating antibody in an ELISPOT system

Reconstitution: Dissolve the contents of one vial by injection of 0.25 ml distilled water into the vial and dilute 100 times in PBS. The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; 50 µl/well).

Detection antibodies (2 vials)

Product: Biotinylated polyclonal antibody to monkey interleukin 12 (IL-12)

Isotype: Rabbit IgG

Purification: Ammonium sulphate precipitation and protein A-affinity chromatography

Labeling: With Biotin-7-NHS (N-hydroxysuccinimide)

Contents: Each vial contains sufficient material for five 96-well ELISPOT plates

Buffer: Prior to lyophilization: 0.5 ml PBS + 1% BSA + 125 mM trehalose

Application: Detection antibody in an ELISPOT system

Reconstitution: Dissolve the contents of one vial by injection of 0.5 ml distilled water into the vial and dilute 100 times in Dilution buffer (see Technical Data Sheet). The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; 100 µl/well).

General

Specificity: Validated for detecting rhesus macaque, cynomolgus monkey, pig-tailed macaque and baboon IL-12

Sterility: Membrane filtered (0.2 µm)

Stability: The lyophilized products are stable for at least one year at 4°C (expiry date is indicated on the vials).
After reconstitution, the antibodies are stable for several months at 4°C (if kept sterile) or for minimal one year at -20°C.