


Typical data Human TNF- α ELISA kit

Catalogue no.: CT209A (5-plate format)

Validated for:

Human



Example of a representative Human TNF- α standard curve using a 4-parameter logistic fit.

Recommended assay range: 2-128 pg/ml.
 Sensitivity: 2 pg/ml.
 Specificity: natural and recombinant human TNF- α .
 Calibration: recombinant human TNF- α .
 Type of sample: serum, plasma and culture supernatant.
 No interference by the sTNF receptors (RI/RII).

References list:

- Chang, W.L.W., Baumgarth, N., Yu, D. and Barry, P.A. 2004. *J. Virol.* 78: 8720-8731
 Chiriva-Interati, M., Mirandola, L., D'Cunha, N., Hardwicke, F., Cannon, M.J., Cobos, E. and Kast, M.W. 2012. *Prostate* 72: 12-23
 Jagessar S.A., Kap Y.S., Heijmans N., van Driel N., van Straalen L., Bajramovic J.J., Brok H.P., Blezer E.L., Bauer J., Laman J.D. and 't Hart B.A. 2010. *J. Neuropathol. Exp. Neurol.* 69: 372-385
 Skov, L., Allen, M.H., Bang, B., Francis, D., Barker, J.N.W.N. and Baadsgaard, O. 2003. *Exp. Dermatol.* 12: 772-776
 Tang, Y-M., Chen, M-H., Chen, G-H., Cai, C-J., He, X-S., Lu, M-G. and Bao, W-M. 2005. *World J. Gastroenterol.* 11: 4574-4578
 Weinhold, M., Eisenblätter, M., Jasny, E., Fehlings, M., Finke, A., Gayum, H., Rüschen-dorf, U., Renner Viveros, P., Moos, V., Allers, K., Schneider, T., Schaible, U.E., Schumann, R.R., Mielke, M.E. and Ignatius, R. 2013. *PLoS ONE* 8: e65934

More references can be found in our References Database on our website: www.ucytech.com/references.